



# Specification

---

1.05739.0000 Molecular sieve 0.4 nm beads, with moisture indicator ~ 2 mm / ~ 10 mesh

---

---

## Specification

---

Identity	passes test	
Loss on ignition (950 °C)	≤ 5	%
Water absorption capacity (24 hrs., 80 % relative humidity)	≥ 20	%

Julian Hofmann

Responsible laboratory manager quality control

This document has been produced electronically and is valid without a signature.